

Aiken - CLOSE

1/30/03

Aiken

@

JG, KC

~ 13:30

wx - 3% cc, cumulus; wind 10 mph E

photo on Karen's camera

- no survey for end of season

control looks fine, probes too

No Flow

10 channels

Stage	1	1.01
WT	2	8.3
SC	3	-9.3
Volts	4	13.76

S/m's creio settings o.k.

changed storage module @ 13:23

took out FSA put in F5B

N₂ Pressure 1500/10 → did not change tank

has full tank in Box