

LEVEL NOTES

STREAM Anderson Creek  
 LOCALITY \_\_\_\_\_  
 PARTY ANW + TJK DATE 2011-11-27, 19\_\_

STATION	B. S.	HT. INST.	F. S.	ELEVATION	REMARKS
					arrive @ box 1220 w/TJK, box is clear of snow and pit.
					- control is filled w/snow and ice
					- ANW will replace old N <sub>2</sub> w/new N <sub>2</sub>
					- new psi @ 1900, system is bubbling o.k.
					- small leak in tank-regulator interface we couldn't essuage w/seal tape... - we'll fix this @ next visit.
					- checking data logger time.
					year 2012 day 335 time 16:41
					changing time from 0000 to 1641
					#6 @ 1642
					- new SM was inserted and
					new program - H1-1213-4.dld
					uploaded to CRUX and
					transferred to SM
					- SM on fill n' stop? yes! ✓
					- data collection interval on 15 min.

