

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 11/14, 1998

MISCELLANEOUS FIELD NOTES

Commonwealth Stream at CI MNG, AB

Opened Gage - ~~Replaced N₂ Cylinder~~
Did NOT Replace N₂ Cylinder,
Old one still had 1100 psi.
kept new tank in box.

Battery 12.86V

Swapped Storage Module
Reset CR10 time

No flow

Control OK, snowcovered.