

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date

11/11

, 19

96

JULIAN 315

MISCELLANEOUS FIELD NOTES

F8 CRESCENT ST

- N₂ DOA
- Batt. 2.7 V on arrival
- Replaced Batt. 12.31 V
- Installed new SM, placed in Batt. 7128"
- Addr. 1 → SM
- Fill & STOP
- SM Batt. ok
- 2 PROG. LOADED
- Set time
- Scan Rate = 900
- * 0 → Log 1
- New N₂ → 1500 PSI NO LEAKS
- Hooked up solar
- Control pretty clean - a little packed snow w/ ice underneath over center of stream.