

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 12-9, 1997

MISCELLANEOUS FIELD NOTES

Crescent Cr @ FB

MDL PJ

1410

sunny cold

$\approx -5^{\circ}\text{C}$

Control has no flow + no ice
NO Evidence of Past Flow

N₂ tank 1900 Psi

Feed 10 Psi

CR10 line is fine

*6 1 -99999

2 7.89

3 0.00

4 13.8



Checked wires - No change

PUE in P20 + checked CI

with volt meter - seems OK.

changed wiring to C2 - no change

next thing is to change pressure
transducer?

Ran levels