

FB - Crescent - OPEN!

17/11/02

weather: 90% overcast

John, Karen

NOTES

No flow, still snow in channel $< 10 \text{ mph}$

Sting module settings

John visited 12/7/02 - see his NBs

N₂ tank - 2000PSI

Reconnected Sting module to blue cable.

Time	Old	New
Julian Day	145	→ 346
Time & yr	OK	

Ch. 1 - Stage = 20.35

Ch. 2 - WT = 6.4225

Ch. 3 - Sp Cond: = 22.1

Ch. 4 - Vol-1.5 = 13.86

Changed scan rate back to 900 sec

✗

tried to Purged outflow line via canoflow

Closed N₂ tank

Then purged line

Air pressures did not drop

~~the~~ - on tank & regulator

→ indicates frozen outflow line