

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 12/18/07, 19

MISCELLANEOUS FIELD NOTES

F8 Crescent

NRM, AMS, LFS

1150 We have flow

batteries are @ 4V

possible problem w/ solar panel / voltage
replacing batteries regulator?

took measurement w/ portable flume

1200 flume .11

tapedown 1.49

N₂ - 900 psi

feed 10 psi

> old batteries were not holding a charge
solar panel and regulator working

New BV - 12.1

1244 1244 *6 - 9999 06 tapedown - 1.33

wt 9.36

cond. - 111

sc 186.67

pH - 7.3

BV 12.24