

U. S. DEPARTMENT OF THE INTERIOR

Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 11/12/98 , 19 98 1315

MISCELLANEOUS FIELD NOTES

F10 - Delta Stream

AB, MG

- Batteries @ 13.5 v each, red light is on!
- cold + snowing!
- replaced N₂ tank high press = 1800 psi
low = 10psi
- swapped SM
- reset CR10 time

NO FLOW, there is minimal ice in the control. No flume here at control nor portable flume.

Build up sand bag walls later this season?