

STATION NUMBER

STREAM House Creek @ 17

PARTY PC, CT

DATE 3 Dec 00, 19

NO. _____ OF _____ SHEETS _____ COMP. BY _____ CHK. BY _____

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 3 Dec 00, 19

mostly clear
pretty warm
20
slight breeze

MISCELLANEOUS FIELD NOTES

House Creek @ H1

PC, CT

- low flow, flow started either 1 Dec or 30 Nov
- N empty - read @ 1000 last time
regulator may have been stuck?
- replacing tanks @ 1325 hrs
- flume running fine
- Still some leakage on left side of control
maybe 10-20' of flow

	inside	outside	
* Stave	2.02	0.15	@ 1328
WT	1.03	0.4°C	
Sc	70.4	31.2 us	
Volts	13.7		

- collected stream chem / fan levels
- Batteries good but solar light not on
- * no bubbles coming out of orifices at flume
may be frozen -