

Site Visit Sheet

Huey Stream

9 Jan. 2013

arrive @ 1820 S.D. and D.C.

weather Mostly cloudy w/ 5.70 clear; slight wind

Conditions Flowing; 2 foot pools in front of control

Found @	*G	Time	O.G.	Time
St.	0.90	1815	0.93	1815
WT	3.5	"	2.6	"
S.Cond	138	"	177	"

* Clean TK probe

WT	2.9	18:30	2.6	18:30
S.Cond.	220		189	

* Son finds CR10 is reading
Conductance Not. S. Cond.

∴ Changed Program

WT	2.4	18:50	2.5	18:50
S. Cond.	198	"	189	

* Son modifies control to make one
pool

* D.C. collects WQ at 19:00

$B_{ski} = 0.09 = 0.0342 cfs$ at 1905

when outside gage = 0.90