

Meas. plots _____ % diff. from rating no. _____ Indicated shift _____

Correct _____

Cross section: _____

Control: _____

Sheet No. _____ of _____ sheets

S 77° 36' 17.9"
E 163° 07' 38.1"

F2 - Huey OK

12/3/02
Tim, Karen

Notes:

No flow

Full of snow

N₂ is empty

GPS works

77° 36' 17.9"
E 163° 07' 38.1"

Taking out F2A storage module

- when attached to keypad giving 100%

Replacing w/ F2B storage module

F2B module - battery installed 3/00

∴ OK!

Has a CR10X

Battery #1 12.6

Battery #2 12.6

New N₂ tank 1700 PSI

NOTE: Extra battery in gauge box

To Do Next Time

Replace F2A string module
no battery date

Ask Chris - why is there an extra
canoflow in F2 box

Ask Chris - if extra panels in gauge
box do they need to be retored

Inventory

distice line

Extra canoflow