

U. S. DEPARTMENT OF THE INTERIOR  
Geological Survey  
WATER RESOURCES DIVISION

STATION NUMBER \_\_\_\_\_

LEVEL NOTES

STREAM \_\_\_\_\_

LOCALITY HUEY CRK

PARTY REW, LFS, ESG DATE 12/7/06, 19\_\_\_\_

STATION	B. S.	HT. INST.	F. S.	ELEVATION	REMARKS
					• removed N-gas bubbling system and replaced with new setup to repair chronic leaking.
- NO FLOW		time = 1515			
<del>A6</del>		1) 2.06			• replaced N tank (pressure ~ 400 psi) Pressure = 1800 psi
					• conoflow appears to be working, but orifice line may still be plugged.
					• orifice frozen