

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

Form 9-275-D
(Jan. 1988)

WATER RESOURCES DIVISION

Date 12-18-03, 19

MISCELLANEOUS FIELD NOTES

Alley Creek CRZ G, NY

- Just Flowing
- Possible N_2 leak, N_2 @ 1400 PSI
- Conflow bubbling too fast
- Tube to transducer really full of fluid, need to try and replace, must find that leak!

#60 1020

stage 1.02
AT -99
WT -99
volts 14.0