

Form 9-276-D  
(Jan. 1988)

U. S. DEPARTMENT OF THE INTERIOR  
Geological Survey

None

WATER RESOURCES DIVISION

Date Nov 14, 19 95 @ 13.15

Kanda

MISCELLANEOUS FIELD NOTES

J.U. + J.S

- Gagebox in good condition
- Old Battery = 11.68V
- Backup: 12.0V → we take battery back
- D/L to sm
- Data logger: 5.38 minutes too slow
- N<sub>2</sub> = 1890 psi
- New Battery: 12.57V

No flow

- Reset data logger
- SM BATT OK
- N<sub>2</sub> new = 1890 psi, no leakage
- bubble rate = 54/min
- Reset sm & time @ 1334