

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

WATER RESOURCES DIVISION

Date 11/10

19 96

JULIAN 314

MISCELLANEOUS FIELD NOTES

F6 SITE

- N₂ TURNED OFF ON ARRIVAL - 1500 psi left
- Batt. unplugged on arrival
- Put in new storage module, powered up "7/28"
- New SM ADDR. = 1
- Fill & Stop = yes
- SM BATT = OK
- 2 programs in SM.
- Set time
- Scan Rate = 900
- * 0
- Hooked up solar panel
- Turned on N₂
- Control - hard snow packed w/ice.