

U. S. DEPARTMENT OF THE INTERIOR
Geological Survey

WATER RESOURCES DIVISION

Date 11/20/99, 19

MISCELLANEOUS FIELD NOTES

Van Gurnard Stn

MG

11/18/99 - changed SM out and used the new one to try the "reset, set to ~~fill~~ fill and stop and copy the program from CR10 to SM" technique seems to have worked, I think.

- CR 10 time was reset on 11/18, it is okay still (@ 1238)
- stage = 0.564
- AT = 0.7
- WT = -7.7 (F6B-14250)
- SC = NONE
- Volt = 14.6

- No flow yet
- flow covered in snow, both orifice and probes under snow.
- swapped N₂ tank