

Date

12 Dec 13 @ 1215

MISCELLANEOUS FIELD NOTES

C1 Canada AM, SWC

Here for telemetry issues, LV's, + Q
flow!Little N_2 leak, fixed. ^{psi @ 1350} box has empty N_2 cylinder!
telemetry: CD (red solid) TX (nothing) ETSL (blink red)

@1219 I6 06

stage 1.42 0.49 poor due to baski

WT 7.57

SC 18.4

baski 0.405 missing ~10% of flow

Telemetry: able to find it kind of, Com3, error: An exception occurred while trying to read this module. Index was out of range. Must be non-negative & less than the size of the collection.
I push "Setup" And all lights turn green. Then only CD stays green some min later.

@1255 I6 06

I6 06
@1302 stage 1.576 0.58baski ~~1.57~~ 0.53 ~ miss 10%

stage 1.59 0.60

WT 8.46 8.3

SC 16.8 22.8

pH 7.20

No. _____ of _____ sheets