

Date

14 Dec 13' @ 1230

MISCELLANEOUS FIELD NOTES

Miers Outlet

No flow, windy & warm. Ran LVLs

N₂ Leak again. Spare tank, but fixed leak @ 1245 ish.
1100 psi, 10, 60 BPM set to.

stage: 0.007 ??? control has 1-2 ft of snow/water/ice
I increase bubble rate & sty spikes, but then goes back to
0.007... Leak in transducer or orifice line? Can't find.

Christ, Orifice to Connoflow leak @ compression fitting
between orif. line & Nut... Had to tighten ridiculously hard.
Leak fixed for now...

was 0.007

*6 @ (329 1) 2.491 2) -3.76 3) -8.7 4) 8.9 5) 13.6

PSI of tank @ 1075... we'll see.

No. _____ of _____ sheets